

This page intentionally left blank.

## Appendix A

### Maximum Permissible Exposure (MPE) for the Eye

**Table A1. Point Source MPEs for the Eye for Commonly Used Lasers**

Laser Type	Wavelength (nm)	MPE ( $\text{W}\cdot\text{cm}^{-2}$ )			
		* $t=0.25$ s	$t=10$ s	$t=600$ s	$t=3 \times 10^4$ s
XeCl <sup>a</sup>	308	—	—	—	$1.33 \times 10^{-6}$
XeF <sup>a</sup>	351	—	—	—	$3.33 \times 10^{-5}$
Argon	514	$2.6 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$
Krypton	530	$2.6 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$
	568	$2.6 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$
	647	$2.6 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$
HeNe	633	$2.6 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$
InGaAlP (diode)	670	$2.6 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$	$1.0 \times 10^{-3}$
GaAs (diode)	808	—	$1.6 \times 10^{-3}$	—	$1.6 \times 10^{-3}$
GaAs (diode)	840	—	$1.9 \times 10^{-3}$	—	$1.9 \times 10^{-3}$
InGaAs (diode)	940	—	$3.0 \times 10^{-3}$	—	$3.0 \times 10^{-3}$
	980	—	$3.6 \times 10^{-3}$	—	$3.6 \times 10^{-3}$
Yb:YAG Q-switched <sup>b</sup>	1030	—	$1.0 \times 10^{-5}$	—	$1.0 \times 10^{-5}$
Nd:YAG Q-switched <sup>b</sup>	1064	—	$2.2 \times 10^{-5}$	—	$2.2 \times 10^{-5}$
Yb: Fiber Q-switched <sup>b</sup>	1070	—	$2.2 \times 10^{-5}$	—	$2.2 \times 10^{-5}$
Yb:YAG (CW)	1030	—	$4.6 \times 10^{-3}$	—	$4.6 \times 10^{-3}$
Nd:YAG (CW)	1064	—	$5.0 \times 10^{-3}$	—	$5.0 \times 10^{-3}$
Yb: Fiber (CW)	1070	—	$5.0 \times 10^{-3}$	—	$5.0 \times 10^{-3}$
CO <sub>2</sub>	10,600	—	0.1	—	0.1

NOTE—This table provides summary point source maximum permissible exposure values for a select group of lasers and key exposure durations. Reference ANSI Z136.1 for comprehensive information.

\*  $t$  is the exposure duration in seconds (s)

<sup>a</sup> When repeated exposure levels are anticipated over two successive days the MPE must be reduced by a factor of 2.5

<sup>b</sup> Operating in repetitive pulsed mode at 11 Hz, 12-ns pulse, 20 mJ/pulse

The information contained in this appendix is not part of this American National Standard (ANS) and has not been processed in accordance with ANSI's requirements for an ANS. As such, this appendix may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to this standard. It is intended for information only.

**Table A2. MPE for the Eye for Selected Single Pulse Lasers**

<b>Laser Type</b>	<b>Wavelength (nm)</b>	<b>Pulse Duration (s)</b>	<b>MPE (J cm<sup>-2</sup>)</b>
Excimer (ArF)	193	$2 \times 10^{-8}$	$3.0 \times 10^{-3}$
Excimer (XeCl)	308	$2 \times 10^{-8}$	$6.7 \times 10^{-3}$
Ruby	694	$1 \times 10^{-3}$	$1.0 \times 10^{-5}$
Yb:YAG (pulsed)	1030	$1 \times 10^{-3}$	$4.6 \times 10^{-5}$
Nd:YAG (pulsed)	1064	$1 \times 10^{-3}$	$5.1 \times 10^{-5}$
Yb:Fiber (pulsed)	1070	$1 \times 10^{-3}$	$5.1 \times 10^{-5}$
Yb:YAG (Q-switched)	1030	$5-100 \times 10^{-9}$	$9.1 \times 10^{-7}$
Nd:YAG (Q-Switched)	1064	$5-100 \times 10^{-9}$	$2.0 \times 10^{-6}$
Yb:Fiber (Q-switched)	1070	$5-100 \times 10^{-9}$	$2.0 \times 10^{-6}$
Carbon Dioxide (CO <sub>2</sub> )	10,600	$1 \times 10^{-3}$	0.1
NOTE—Table A2 is provided as a reference of MPEs for a selected set of single pulse lasers for the noted wavelengths and duration levels.			

**Table A3. Typical Laser Hazard Classification, Continuous Wave (CW) Lasers**

Laser Type	Wavelength (nm)	Power Range (W)	Maximum Hazard Class
Helium-Cadmium (HeCd)	325	$1.6 \times 10^{-5}$ to 0.5	Class 3B
Argon (Ar)	351, 363	$1.6 \times 10^{-5}$ to 0.5	Class 3B
Argon (Ar)	488, 514	$\leq 0.001$	Class 2
Helium-Neon (HeNe)	633	$\leq 0.001$	Class 2
InGaAlP (diode)	670	$\leq 0.001$	Class 2
Argon (Ar)	488, 514	0.001 to 0.005	Class 3R
Helium-Neon (HeNe)	633	0.001 to 0.005	Class 3R
InGaAlP (diode)	670	0.001 to 0.005	Class 3R
Argon (Ar)	488, 514	$> 0.5$	Class 4
InGaAlP (diode)	670	$> 0.5$	Class 4
Dye	400 to 550	$> 0.5$	Class 4
InGaAs (diode)	940, 980	$> 0.5$	Class 4
GaAlAs (diode)	780	0.0028 to 0.5	Class 3B
Yb:YAG	1030	0.0088 to 0.5	Class 3B
Nd:YAG	1064	0.0096 to 0.5	Class 3B
Yb:Fiber	1070	0.0096 to 0.5	Class 3B
InGaAsP (diode)	1310	0.148 to 0.5	Class 3B
Doubled Yb:YAG	515	$> 0.5$	Class 4
Doubled Nd:YAG	532	$> 0.5$	Class 4
Doubled Yb:Fiber	535	$> 0.5$	Class 4
Yb:YAG	1030	$> 0.5$	Class 4
Nd:YAG	1064	$> 0.5$	Class 4
Yb:Fiber	1070	$> 0.5$	Class 4
Carbon dioxide (CO <sub>2</sub> )	10,600	$> 0.5$	Class 4

NOTE—Table A3 provides a quick reference for some of the various lasing mediums, lasers and laser systems that may find use in educational institutions. For a respective lasing medium and wavelength, its corresponding hazard classification is provided for an output emission level, based upon an eight-hour exposure duration (most restrictive). Those wavelengths below 400 nm do not include the reduction factor for two successive exposure eight-hour shifts within a 24-hour period per Z136.1-2014, Section 8.

This table is not to serve for reclassification purposes. It is only for summary guidance as to the applicable hazard class associated with a laser and its output characteristics. The appropriate documentation and assessment for the subject laser or laser system shall prevail. Reference ANSI Z136.1 (latest revision) for comprehensive information pertaining to laser hazard classification

#### A4. Bibliography.

Laser Institute of America. (2014). *American national standard for safe use of lasers* (ANSI Z136.1-2014).

Mathes, R., Feychting, M., Ahlborn, A., Breitbart, E., Croft, R., Green, A., de Grujil, F.R., Hietanen, M., Jokela, K., Lin, J.C., Marino, C., Peralta, AP., Saunders, R., Schulmeister, K., Sienkiewicz, Z., Söderberg, P., Stuck, B.E., Swerdlow, A.J., Taki, M., van Rongen, E.,...Zuchlich, J.A. (2013). ICNIRP guidelines on limits of exposure to laser radiation of wavelengths between 180 nm and 1,000  $\mu\text{m}$ . *Health Physics Society*, 105(3): 271-295. <https://doi.org/10.1097/HP.0b013e3182983fd4>

## Appendix B

### Control Measures for Laser Classes

**Table B1. Engineering Controls**

Engineering Control Measures		Classification						
		1	1M	2	2M	3R	3B	4
Protective Housing (4.5.1, 4.5.1.2)		X	X	X	X	X	X	X
Without Protective Housing (4.5.1.1)		LSO shall establish Alternative Controls						
Interlocks on Removable Protective Housings (4.5.2)		▽	▽	▽	▽	▽	X	X
Service Access Panel (4.5.3)		▽	▽	▽	▽	▽	X	X
Key Control (4.5.4)		—	—	—	—	—	•	X
Viewing Windows and Display Screens (4.5.5.1)		Assure viewing limited < MPE						
Collecting Optics (4.5.5.2)								
Fully Open Beam Path (4.5.6.1)		—	—	—	—	—	X NHZ	X NHZ
Limited Open Beam Path (4.5.6.2)		—	—	—	—	—	X NHZ	X NHZ
Enclosed Beam Path (4.5.6.3)		None is required if 4.3.1 and 4.3.2 fulfilled						
Remote Interlock Connector (4.5.7)		—	—	—	—	—	•	X
Beam Stop or Attenuator (4.5.8)		—	—	—	—	—	•	X
Legend:	X Shall	▽ Shall, if enclosed Class 3B or 4						
	• Should	MPE Shall, if MPE is exceeded						
	— No requirement	NHZ Nominal Hazard Zone analysis required						
		* May apply with use of optical aids						

(Table B1 Continued on Next Page)

**Table B1. Engineering Controls (cont.)**

Engineering Control Measures		Classification						
		1	1M	2	2M	3R	3B	4
Activation Warning Systems (4.5.9)		—	—	—	—	—	•	X
Indoor Laser-Controlled Area (4.5.10)		—	*	—	*	—	X NHZ	X NHZ
Class 3B Indoor Laser-Controlled Area (4.5.10.1)		—	—	—	—	—	X	—
Class 4 Laser-Controlled Area (4.5.10.1, 4.5.10.2)		—	—	—	—	—	—	X
Outdoor Control Measures (4.5.11)		X	* NHZ	X NHZ	* NHZ	X NHZ	X NHZ	X NHZ
Laser in Navigable Airspace (4.5.11.1)		X	* NHZ	X NHZ	* NHZ	X NHZ	X NHZ	X NHZ
Temporary Laser-Controlled Area (4.5.12)		∇ MPE	∇ MPE	∇ MPE	∇ MPE	∇ MPE	—	—
Remote Firing and Monitoring (4.5.13)		—	—	—	—	—	—	•
Equipment Labels (4.5.14)		X	X	X	X	X	X	X
Laser Area Warning Signs and Activation Warnings (4.5.15)		—	—	—	—	•	X NHZ	X NHZ
Legend:		X	Shall	∇	Shall, if enclosed Class 3B or 4			
		•	Should	MPE	Shall, if MPE is exceeded			
		—	No requirement	NHZ	Nominal Hazard Zone analysis required			
				*	May apply with use of optical aids			

**Table B2. Administrative and Procedural Controls**

Administrative and Procedural Control Measures			Classification						
			1	1M	2	2M	3R	3B	4
Standard Operating Procedures (4.6.1)			—	—	—	—	—	•	X
Output Emission Limitations (4.6.2)			—	—	—	—	LSO Determination		
Education and Training (4.6.3)			—	•	•	•	•	X	X
Authorized Personnel (4.6.4)			—	*	—	*	—	X	X
Alignment Procedures (4.6.5)			∇	∇	∇	∇	∇	X	X
Protective Equipment (4.6.6)			—	*	—	*	—	•	X
Spectators (4.6.7)			—	*	—	*	—	•	X
Service Personnel (4.6.8)			∇	∇	∇	∇	∇	X	X
Laser Optical Fiber Transmission Systems			MPE	MPE	MPE	MPE	MPE	X	X
Laser Robotic Installations (4.5.16)			—	—	—	—	—	X NHZ	X NHZ
Protective Eyewear (4.6.6.2)			—	—	—	—	—	•	X
Window Protection (4.6.6.3)			—	—	—	—	—	X NHZ	X NHZ
Protective Barriers and Curtains (4.6.6.1)			—	—	—	—	—	•	•
Skin Protection (4.6.6.5)			—	—	—	—	—	X MPE	X MPE
Other Protective Equipment (4.6.9)			Use may be required						
Warning Signs and Labels (Design Requirements) (4.5.14, 4.5.15)			—	—	•	•	•	X NHZ	X NHZ
Service Personnel (4.6.8)			LSO Determination						
Laser System Modifications (4.1.3)									
Legend:	X	Shall	∇	Shall, if enclosed Class 3B or 4					
	•	Should	MPE	Shall, if MPE is exceeded					
	—	No requirement	NHZ	Nominal Hazard Zone analysis required					
			*	May apply with use of optical aids					

### B3. Bibliography.

Laser Institute of America. (2014). *American national standard for safe use of lasers* (ANSI Z136.1-2014).



## Appendix C

### Laser Beam Hazard Terms and Assessment for Selected Lasers

#### C1. General

The following is provided as summary information for laser hazard assessment and key terms of use. Reference ANSI Z136.1 (latest revision) for more comprehensive information.

Many examples available will illustrate the eye as the part of the human body exposed, as this is generally the most sensitive organ at risk. Hazard analysis must account for eye and skin exposure conditions, and then apply the most restrictive evaluation for establishing the hazard areas and respective safety control measures.

#### C2. Laser Beam Profile and Dimensions

Beam profiles are referenced against an idealized Gaussian beam. For laser safety calculations, the beam diameter and beam divergence values are measured at the 1/e points of the beam intensity profile. This is the radial position at which the beam intensity has dropped to 1/e (~37%) of its peak value.

Using this for the beam diameter will then provide the peak intensity values when using beam power or energy in the respective equations:

$$\text{Irradiance} = \text{Power intensity} = \frac{\text{Power}}{\text{Area}} = \frac{\text{Power}}{\pi(r_{(1/e)})^2} = \frac{4 \cdot \text{Power}}{\pi(d_{(1/e)})^2} \quad (\text{C1})$$

$$\text{Radiant exposure} = \text{Energy intensity} = \frac{\text{Energy}}{\text{Area}} = \frac{\text{Energy}}{\pi(r_{(1/e)})^2} = \frac{4 \cdot \text{Energy}}{\pi(d_{(1/e)})^2} \quad (\text{C2})$$

Where,

$r_{(1/e)}$  = beam radius at 1/e intensity point,

$d_{(1/e)}$  = beam diameter at 1/e intensity points, and

$a$  = beam diameter (1/e points) at the laser exit port within the Z136 family of standards.

In many cases, the manufacturer may provide the 1/e<sup>2</sup> beam diameter information. Using this for intensity calculations will provide only the average intensity value for the beam.

For laser safety analysis and calculations, it is the peak intensity that is of interest.

To determine the  $d(1/e)$  value from a given  $d(1/e^2)$  beam diameter, then,

$$d(1/e) = \frac{d(1/e^2)}{\sqrt{2}} \quad (C3)$$

Figure C1 illustrates this relationship.

Using Z136 notation, then, at the beam exit port:

$$\text{Irradiance}_{\text{exit port}} = E_0 = \frac{\text{Power}}{\text{Area}} = \frac{\text{Power}}{\pi(r_{(1/e)})^2} = \frac{4 \cdot \text{Power}}{\pi(d_{(1/e)})^2} = \frac{4 \cdot \Phi}{\pi a^2} \quad (C4)$$

$$\text{Radiant exposure}_{\text{exit port}} = H_0 = \frac{\text{Energy}}{\text{Area}} = \frac{\text{Energy}}{\pi(r_{(1/e)})^2} = \frac{4 \cdot \text{Energy}}{\pi(d_{(1/e)})^2} = \frac{4 \cdot Q}{\pi a^2} \quad (C5)$$

Where,

$E_0$ =irradiance, in terms of Watt per square centimeter,  $\text{W}/\text{cm}^2 = \text{W} \cdot \text{cm}^{-2}$ ,

$H_0$  = radiant exposure, in terms of Joules per square centimeter,  $\text{J}/\text{cm}^2 = \text{W} \cdot \text{cm}^{-2}$ ,

$a$  = beam diameter (1/e points) at the laser exit port, in terms of centimeters (cm),

$\Phi$  = power, in terms of Watts (W), and

$Q$  = energy, in terms of Joules (J).

### C3. Types of Exposure Conditions

There are typically three types of exposure conditions referenced when dealing with laser safety. The illustrations provided are with a collimated beam arrangement, although other exposure conditions with different beam arrangements can also apply. If it is not reasonably foreseeable to be dealing with a given exposure condition for a specific lesson plan or experiment, then the associated hazard distance will not apply.

**C3.1 Intrabeam Viewing.** This is direct exposure within the beam. With a collimated beam generally, this will produce the greatest hazard distance (see Figure C2).

**C3.2 Specular Reflection.** This is a mirror like reflection where the redirected beam essentially maintains its original properties. The total hazard distance will be essentially the same as that for an intrabeam condition (see Figure C3).